



331 Treble Cove Road
North Billerica, MA 01862

800.362.2668
www.lantheus.com

March 14, 2016

**PLEASE READ CAREFULLY FOR IMPORTANT SAFETY INFORMATION ABOUT
XENON-133 (Xenon Xe 133 Gas) DOSING**

To Our Valued Nuclear Medicine Customers:

We wish to provide you today with information regarding the only safe and FDA-approved use of our Xenon-133 Calidose™ Dispenser System and our Xenon-133 (Xenon Xe 133 Gas) vials.

Our Calidose™ Dispenser System is specifically designed and approved by the FDA for use with Lantheus' Xenon-133 (Xenon Xe 133 Gas) vials only, as stated in our Prescribing Information and instructions for use (attached). **Using Xenon-133 vials of a different size or shape with our Calidose™ Dispenser System has not been proven to be safe.**

By following our time-tested, FDA-approved instructions, you will minimize the risk of radiation exposure for your patients and your clinical team.

As always, Lantheus Medical Imaging appreciates your support and thanks you for your business. If you have any questions with regards to the foregoing, or need any assistance, please do not hesitate to contact our Medical Affairs team at 978-667-9531.

Sincerely,

A handwritten signature in blue ink, appearing to read "Ces Orlandi", with a stylized flourish at the end.

Ces Orlandi, M.D.
Chief Medical Officer